

A Mass-Flow-Calorimetry System for Scaled-up Experiments on Anomalous Heat Evolution at Elevated Temperatures

A. Kitamura^{1,2}, A. Takahashi^{1,3}, R. Seto¹, Y. Fujita¹, A. Taniike² and Y. Furuyama²

¹ Technova Inc., Japan, kitamuraakira3@gmail.com

² Kobe University, Japan, ³ Osaka University, Japan

We have been studying phenomena of anomalous heat evolution from hydrogen-isotope-loaded nano-composite samples at elevated temperatures as well as at room temperature using a twin absorption system [1, 2]. Recent experiments have used Ni-based nano-composite samples; Pd₁Ni₇/ZrO₂ (“PNZ”), Ni/ZrO₂ (“NZ”), Cu_{0.081}Ni_{0.36}/ZrO₂ (“CNZ”) and Cu_{0.21}Ni_{0.21}/ZrO₂ (“CNZII”). The results of measurements have been presented in the meetings of the 12th Japan CF-Research Society (JCF12), the 17th International Conference on Condensed Matter Nuclear Science (ICCF17) and the 13th Japan CF-Research Society (JCF13), and have been/will be published in [3], [4] and [5], respectively.

These will be summarized, and the time-dependent data will be re-analyzed in another paper by A. Takahashi in this Conference for speculating heat releasing mechanisms during the several-week-lasting phase of D(H)-loading into the nano-composite samples. As will be shown there, a lot of interesting, even astonishing, features are involved; burst-like heat release with anomalously high values of differential heat of sorption (η) reaching ca. 600 eV/atom-H, large values of integrated heat reaching ca. 800 eV/atom-Ni from the CNZ sample absorbing H, and abrupt desorption with absorbed energy of 50 - 80 eV/atom-Ni observed almost exclusively in the first 573-K run for each sample.

To confirm the interesting phenomena, repeated measurements with improved signal-to-noise ratio are required. Since the easiest way for this is to increase the sample amount, we have fabricated a reaction chamber with a ten-times-larger volume than in-being one. Another important improvement is a mass flow calorimetry applied to the system using an oil coolant with a boiling point of 390 deg-C. Moreover, to make residual gas mass spectral analysis in $A = 1 - 6$ amu range, a QMA system is going to be installed in the line of the apparatus.

In the presentation we will show the schematics of this new oil-cooling mass-flow calorimetry system for observing anomalous heat evolution in H(D)-gas charging to Ni-based nano-composite samples and for calibration runs using blank alumina sample.

- [1] Akira Kitamura, Yuki Miyoshi, Akira Taniike, Akito Takahashi, Reiko Seto and Yushi Fujita; J. Condensed Matter Nucl. Sci. **4** (2011) 56-68.
- [2] Y. Miyoshi, H. Sakoh, A. Taniike, A. Kitamura, A. Takahashi, R. Seto and Y. Fujita; J. Condensed Matter Nucl. Sci. **10** (2013) 46-62.
- [3] Y. Miyoshi, H. Sakoh, A. Taniike, A. Kitamura, A. Takahashi, R. Seto and Y. Fujita; Proc. JCF12 (2012) 1-9.
- [4] H. Sakoh, Y. Miyoshi, A. Taniike, Y. Furuyama, A. Kitamura, A. Takahashi, R. Seto, Y. Fujita, T. Murota, T. Tahara; to be published in Proc. ICCF17.
- [5] H. Sakoh, Y. Miyoshi, A. Taniike, Y. Furuyama, A. Kitamura, A. Takahashi, R. Seto, Y. Fujita, T. Murota, T. Tahara; to be published in Proc. JCF13.